

SUPPLEMENTAL/BID BULLETIN

Republic of the Philippines
Metals Industry Research and Development Center
Gen. Santos Ave., Bicutan, Taguig City 1631
Tel Nos. 837 0431 to 38 loc. 402; Fax No. 838 7876
Email: bac@mirdc.dost.gov.ph

ADDENDUM NO. 1

This Supplemental Bid Bulletin No. 1 is being issued to further clarify, modify and amend items, specifications in the Bid Documents to address key issues for the “**PROCUREMENT OF ONE (1) UNIT DEADWEIGHT PRESSURE TESTER AND ACCESSORIES**”

DEADWEIGHT PRESSURE TESTER	
FROM	TO
OPENING OF BID SCHEDULE	
August 07, 2018; 01:00 P.M.	August 09, 2018; 01:00 P.M.
TERMS OF REFERENCE	
Accessories	
1. Two (2) units digital test gauge needed for the calibration of low and medium ranges of pressure gauges. The range must be within 0 - 35 MPa and 0 – 70 MPa both with an accuracy of 0.02% full scale.	1. Two (2) units digital test gauge needed for the calibration of low and medium ranges of pressure gauges. The range must be within 0 - 35 MPa and 0 – 70 MPa both with an accuracy of 0.02% full scale with a manual pressure generator.
2. Duty Workbench with the following specifications: top size at least 1.2 x 0.7 meters, two (2) locking bench drawers with pull handles and ball bearing drawer slides, with built-in 6” jaw bench-vise.	2. Duty Workbench with the following specifications: top size at least 1.2 x 0.7 meters, two (2) locking bench drawers with pull handles and ball bearing drawer slides, material: powder-coated steel, with built-in 6” jaw bench-vise.
General	
Supplier must conduct at least eight (8) hours training on the operation of the unit for at least six (6) personnel.	Supplier must conduct at least eight (8) hours training on the operation of the unit for at least six (6) personnel. All training materials needed shall be shouldered by the supplier.
Delivery period must be within 30 calendar days.	Delivery period must be within 60 calendar days including commissioning, testing and training.

This shall form an integral part of the Bidding Documents for the “**PROCUREMENT OF ONE (1) UNIT DEADWEIGHT PRESSURE TESTER AND ACCESSORIES**”

For the guidance and information of all concerned.

MS. AUREA T. MOTAS
BAC Chairman
Metals Industry Research and Development Center