

SUPPLEMENTAL/BID BULLETIN

Republic of the Philippines
Metals Industry Research and Development Center
Gen. Santos Ave., Bicutan, Taguig City 1631
Tel Nos. 837 0431 to 38 loc. 402; Fax No. 838 7876
Email: bac@mirdc.dost.gov.ph

ADDENDUM NO. 1

This Supplemental Bid Bulletin No. 1 is being issued to further clarify, modify and amend items, specifications in the Bid Documents to address key issues for the **“PROCUREMENT OF ONE (1) LOT PERCHLORIC ACID FUMEHOOD”**

TERMS OF REFERENCE	
FROM	TO
Delivery period is 60 days after receipt of NTP.	Delivery, installation, commissioning, testing and acceptance shall be within 90 calendar days after receipt of NTP.
TECHNICAL SPECIFICATION	
Additional	blower/pump should be corrosion resistant

This shall form an integral part of the Bidding Documents for the **“PROCUREMENT OF ONE (1) LOT PERCHLORIC ACID FUMEHOOD”**

For the guidance and information of all concerned.

Sgd.

MS. AUREA T. MOTAS

BAC Chairman

Metals Industry Research and Development Center